



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R60/2000-NL1-16.15  
Project number SO16201399  
Page 1 of 2

Issuing authority	NMI Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Ningbo ETDZ AIEN Electronics Co.,Ltd No.66, XianTan road, LandXia Street 315480 YuYao city, Ningbo China
Identification of the certified type	A <b>compression load cell</b> , with strain gauges. Type : IN-ZS
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R60** - Edition 2000 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMI Certin B.V., OIML Issuing Authority NL1**  
2 May 2016

  
C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see [www.nmi.nl](http://www.nmi.nl)).

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R60/2000-NL1-16.15  
Project number SO16201399  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-13200577-01 dated 10 November 2015 that includes 51 pages.

## Characteristics of the load cell:

Maximum capacity ( $E_{max}$ )	10 t up to and including 50 t
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	2,0 mV/V
Maximum number of load cell intervals (n)	4000
Ratio of minimum LC Verification interval	15000
$Y = E_{max} / V_{min}$	4000
Ratio of minimum dead load output return	
$Z = E_{max} / (2 * DR)$	
Input impedance	700 $\Omega \pm 15 \Omega$
Temperature range	-10 °C / +40 °C
Fraction $p_{LC}$	0,7
Humidity Class	CH
Safe overload	150 % of $E_{max}$
Output impedance	700 $\Omega \pm 5 \Omega$
Recommended excitation	10 V AC / DC
Excitation maximum	15 V AC / DC
Transducer material	Steel
Atmospheric protection	Hermetically welded

The characteristics for  $n_{max}$  and Y can be reduced separately. Z is proportional or equal to  $n_{max}$ .

Each produced load cell is provided with an accompanying document with information about its characteristics.

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by the United States of America (NIST Handbook 44 and NCFM Publication 14), included in the MAA Declaration of Mutual Confidence:

- R 60 DoMC-01 rev.0, Additional requirements from the United States;
- R 60 DoMC-02 rev.0, Additional requirements from the United States.



TRUE VALUE

Ningbo AIEN Electronics Co.,Ltd  
Attn. Mr. Edward Zha  
No.66 Xian Tan Road ,LangXia Street  
Yuyao City, Ningbo 315480  
P.R. China

Date	Your reference	Our reference
14 July 2016	-	CPC-SO16201399-SBa-AB

Subject  
OIML MAA CoC – project closure

Dear Mr. Zha,

With this letter we send you the following document:

- The parallel OIML MAA Certificate of Conformity, number R60/2000-NL1-16.15.

Hereby this project is considered to be closed. If you have any questions, please don't hesitate to contact us by email at [weighing@nmi.nl](mailto:weighing@nmi.nl).

With kind regards,  
NMI Certin B.V.

Mrs. Sandra Baan  
Administration Product Certification

NMI Certin B.V.  
Hugo de Grootplein 1, 3314 EG Dordrecht  
P.O. Box 394, 3300 AJ Dordrecht, Netherlands  
T +31 78 6332 332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
<http://nmi.nl>

NMI Certin B.V., chamber o.c. no. 27.233.418